

CERTIFICATE OF ANALYSIS

Product: Quintus 5-Part Diluent

Lot No: 2206-201

Product Part #: 501-292

Exp. Date: 2024-05-31

Manufacturing Date: 2022-06-24


Quality Testing Date: 2022-06-28

Parameter	Specification	Result	Pass/Fail
pH (units):	6.7 – 7.1	7.1	Pass
Osmolality (mOsm/kg):	320 – 350	328	Pass
Conductivity (mS/cm):	18.50 – 19.50	18.72	Pass
Background (x10 ³ /μL)	0 – 5	3	Pass
Functional Test	Pass	Pass	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.



Jeanne Roscoe
Quality and Regulatory Manager



Date

