

Certificate of Analysis

Product MEDONIC M51-L1 LYSE (200 mL)
Product No 1504511
Lot No 2022061601
Manufacturing Date 2022-06-16
Expiration Date 2024-06-15


Test	Specification
pH	2,0 – 2,4
CBC Performance test*	Passes test
Background WBC ($\times 10^9/L$)	$\leq 0,2$

*Performance testing at a reference laboratory. WBC, HGB and WBC scattergrams tested and approved.

Actual values meet specification.

Date: 2022-08-04


Sofia Karlström
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

