


Certificate of Analysis

Product EXIGO DILUENT, RFID
Product No 1504504 (10 L)
Lot No 2208-771
Manufacturing Date 2022-08-17
Expiration Date 2025-08-17


Test	Specification
pH	6.90 – 7.20
Osmolality (mOsm/kg)	300 – 330
Conductivity (mS/cm)	18.00 - 19.00
Particles (PLT/RBC channel)	≤ 5
Total bacterial growth (CFU/100ml)	≤ 5

Actual values meet specifications.

Date: 2022-08-25



Caroline Sparf
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

