

---

# Certificate of Analysis

<b>Product</b>	MEDONIC M-SERIES LYSE, RFID (5 L)
<b>Product No</b>	1504461
<b>Lot No</b>	2208-775
<b>Manufacturing Date</b>	2022-08-22
<b>Expiration Date</b>	2025-08-22

---

<b>Test</b>	<b>Specification</b>
pH	6,80 – 7,20
Osmolality (mOsm/kg)	220 – 240
Conductivity (mS/cm)	14,00 – 15,00
CBC Performance test*	Passes test
Particles (WBC channel)	≤ 0,2

\* Performance test using calibrated Blood cell counter at Boule Medical reference laboratory. All parameters except mid cells are tested. Accepted reference 95 % to 105 % of expected control blood sample values.

Actual value meets specification.

Date: 2022-08-31



*Sofia Karlström*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

