

Certificate of Analysis

Product MEDONIC M-SERIES DUALPACK (1 L Lyse / 5 L Diluent)
Product No 1504128
Lot No 2208-776
Manufacturing Date 2022-08-15
Expiration Date 2025-08-15

Test	Specification Diluent	Specification Lyse
pH	6,90 – 7,10	6,80 – 7,20
Osmolality (mOsm/kg)	290 – 310	220 – 240
Conductivity (mS/cm)	17,50 – 18,00	14,00 – 15,00
CBC Performance *	Passes test	Passes test
Particles (PLT/WBC channels)	≤ 5	≤ 0,2

* Performance test using calibrated Blood cell counter at Boule Medical reference laboratory. All parameters except mid cells are tested. Accepted reference 95 % to 105 % of expected control blood sample values.

Actual values meet specifications.

Date: 2022-08-31



Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

