

## CERTIFICATE OF ANALYSIS

**Product:** Medonic™ M-Series Diluent

**Lot No:** 2206-002

**Product Part #:** 501-211

**Exp. Date:** 2025-05-31

**Manufacturing Date:** 2022-06-06

**Quality Testing Date:** 2022-06-09

Parameter	Specification	Result	Pass/Fail
pH (units):	6.90 – 7.20	7.10	Pass
Osmolality (mOsm/kg):	290 – 310	292	Pass
Conductivity (mS/cm):	17.50 – 18.00	17.51	Pass
Background Test (x10 <sup>3</sup> /μL):	0.0 – 5.0	0.25	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.

  
\_\_\_\_\_  
Jeanne Roscoe  
Quality and Regulatory Manager

6/14/2022  
Date

