

Certificate of Analysis

Product Quintus 5-part Diluent 20L
Product No 1504291
Lot No 2208-777
Manufacturing Date 2022-08-24
Expiration Date 2024-08-24

Test	Specification
pH	6,8 – 7,2
Osmolality (mOsm/kg)	315 – 350
Conductivity (mS/cm)	18,8 – 19,2
Background PLT (x 10 ⁹ /L)	≤ 5

Actual values meet specification.

Date: 2022-09-06



Pernilla Hammarstedt Rydberg
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

