

# Certificate of Analysis

**Product** SWELAB ALFADILUENT, RFID (20 L)  
**Product No** 1504462  
**Lot No** 2208-780  
**Manufacturing Date** 2022-08-28  
**Expiration Date** 2025-08-28

| Test                        | Specification |
|-----------------------------|---------------|
| pH                          | 6,70 – 6,90   |
| Osmolality (mOsm/kg)        | 294 – 312     |
| Conductivity (mS/cm)        | 16,98 – 17,93 |
| Particles (PLT/RBC channel) | ≤ 5           |

Actual value meets specification.

Date: 2022-09-06



*Pernilla Hammarstedt Rydberg*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

