

# Certificate of Analysis

**Product** SWELAB ALFALYSE (5 L)  
**Product No** 1504125  
**Lot No** 2209-785  
**Manufacturing Date** 2022-09-07  
**Expiration Date** 2025-09-07

---

Test	Specification
pH	7,80 – 8,30
Osmolality (mOsm/kg)	295 – 330
Conductivity (mS/cm)	17,75 – 18,25
CBC Performance *	Passes test
Particles (WBC channel)	≤ 0,2

\* Performance test using calibrated Blood cell counter at Boule Medical reference laboratory. All parameters except mid cells are tested. Accepted reference 95 % to 105 % of expected control blood sample values.

Actual value meets specification.

Date: 2022-09-14



*Caroline Sparf*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

