a Boule Diagnostics, AB company



CERTIFICATE OF ANALYSIS

Product: Boule Cleaning Kit Kit Lot No: 1699

Product Part #: 501-036 Kit Exp. Date: 2024-07-31

Part Number: 503-085			
Boule Detergent Cleaner Lot	# 1700		
Manufacturing Date: 2022-07-15		Quality Testing Date: 2022-07-19	
Parameter	Specification	Result	Pass/Fail
pH (units):	7.60 - 8.40	8.29	Pass

Part Number: 503-086					
Boule Hypochlorite (2%) Cleaner Lot # 1701					
Manufacturing Date: 2022-07-12	Quality Testing Date: 2022-07-13				
Parameter	Specification	Result	Pass/Fail		
pH (units):	11.80 - 13.00	12.95	Pass		

Part Number: 503-087			
Boule Enzymatic Cleaner Lot # 1	702		
Manufacturing Date: 2022-07-15	Quality Testing Date: 2022-07-19		
Parameter	Specification	Result	Pass/Fail
pH (units):	7.60 - 8.40	8.09	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.

Jeanne Roscoe

Quality and Regulatory Manager

Date