

## CERTIFICATE OF ANALYSIS

**Product:** Swelab™Alfalyse, 5L

**Lot No:** 2209-003

**Product Part #:** 501-207


**Exp. Date:** 2025-08-31

**Manufacturing Date:** 2022-09-27

**Quality Testing Date:** 2022-09-30

Parameter	Specification	Result	Pass/Fail
pH (units):	7.54 – 8.30	8.28	Pass
Osmolality (mOsm/kg):	295 – 330	316	Pass
Conductivity (mS/cm):	18.00– 18.50	18.30	Pass
Functional Test	Pass	Pass	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.

  
\_\_\_\_\_  
Jeanne Roscoe  
Quality Assurance Manager

10/12/22  
Date

