

CERTIFICATE OF ANALYSIS

Product: Swelab™ AlfaDiluent RFID

Lot No: 2209-004

Product Part #: 501-465

Exp. Date: 2025-08-31

Manufacturing Date: 2022-09-29


Quality Testing Date: 2022-10-03

Parameter	Specification	Result	Pass/Fail
pH (units):	6.50 – 6.90	6.84	Pass
Osmolality (mOsm/kg):	294 – 312	297	Pass
Conductivity (mS/cm):	16.98 – 17.93	17.18	Pass
Background (x10 ³ /μL):	0.0 – 8.0	0.0	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.



Jeanne Roscoe
Quality Assurance Manager



Date

