

Certificate of Analysis

Product	BOULE HYPOCHLORITE 2% CLEANER (500 mL)
Product No	1504113
Lot No	2210-809
Manufacturing date	2022-10-31
Expiration Date	2024-10-31

Test Hypochlorite	Specification
Hypochlorite concentration	2.0 – 2.4 %
pH	11.5 – 12.1
Blank	PLT < 100 x10 ⁹ /L RBC < 0.05 x 10 ¹² /L

Actual values meet specifications.

Date: 2022-11-03



Sofia Karlström
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

