

# Certificate of Analysis

<b>Product</b>	MEDONIC M51-L2 LYSE (500 mL)
<b>Product No</b>	1504512
<b>Lot No</b>	202209041
<b>Manufacturing Date</b>	2022-09-04
<b>Expiration Date</b>	2024-09-03

<b>Test</b>	<b>Specification</b>
pH	4,0 – 4,4
CBC Performance test*	Passes test
Background WBC (x 10 <sup>9</sup> /L)	≤ 0,2

\*Performance testing at a reference laboratory. WBC, HGB and WBC scattergrams tested and approved.

Actual values meet specification.

Date: 2022-10-31



*Sofia Karlström*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

