

# Certificate of Analysis

**Product** MEDONIC M51-L1 LYSE (200 mL)  
**Product No** 1504511  
**Lot No** 2022082501  
**Manufacturing Date** 2022-08-25  
**Expiration Date** 2024-08-24

Test	Specification
pH	2,0 – 2,4
CBC Performance test*	Passes test
Background WBC (x 10 <sup>9</sup> /L)	≤ 0,2

\*Performance testing at a reference laboratory. WBC, HGB and WBC scattergrams tested and approved.

Actual values meet specification.

Date: 2022-10-31



*Sofia Karlström*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

