

# Certificate of Analysis

**Product** SWELAB LUMI-L1 LYSE (200 mL)  
**Product No** 1504515  
**Lot No** 2022082501  
**Manufacturing Date** 2022-08-25  
**Expiration Date** 2024-08-24

| Test                                  | Specification |
|---------------------------------------|---------------|
| pH                                    | 2,0 – 2,4     |
| CBC Performance test*                 | Passes test   |
| Background WBC (x 10 <sup>9</sup> /L) | ≤ 0,2         |

\*Performance testing at a reference laboratory. WBC, HGB and WBC scattergrams tested and approved.

Actual values meet specification.

Date: 2022-10-26



*Sofia Karlström*  
Sofia Karlström  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

