

## CERTIFICATE OF ANALYSIS

**Product:** Quintus 5-Part Diluent

**Lot No:** 2210-201

**Product Part #:** 501-292


**Exp. Date:** 2024-09-30

**Manufacturing Date:** 2022-10-05

**Quality Testing Date:** 2022-11-08

Parameter	Specification	Result	Pass/Fail
pH (units):	6.7 – 7.1	7.1	Pass
Osmolality (mOsm/kg):	320 – 350	328	Pass
Conductivity (mS/cm):	18.50 – 19.50	18.78	Pass
Background ( x10 <sup>3</sup> /μL)	0 – 5	1	Pass
Functional Test	Pass	Pass	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.

  
\_\_\_\_\_  
Jeanne Roscoe  
Quality and Regulatory Manager

11/18/22  
Date

