

CERTIFICATE OF ANALYSIS

Product: Quintus 5-Part Lyse**Lot No:** 2210-301**Product Part #:** 501-291**Exp. Date:** 2024-04-30**Manufacturing Date:** 2022-10-06**Quality Testing Date:** 2022-11-08

Parameter	Specification	Result	Pass/Fail
pH (units):	2.7 – 3.0	3.0	Pass
Osmolality (mOsm/kg):	35 – 44	40	Pass
Conductivity (µS/cm):	1,117 – 1,541	1,221	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.



Jeanne Roscoe
Regulatory and Compliance Manager

11/18/22
Date

