

Certificate of Analysis


Product SWELAB ALFADILUENT (20 L)
Product No 1504124
Lot No 2210-804
Manufacturing Date 2022-10-20
Expiration Date 2025-10-20

Test	Specification
pH	6,70 – 6,90
Osmolality (mOsm/kg)	294 – 312
Conductivity (mS/cm)	16,98 – 17,93
Particles (PLT/RBC channel)	≤ 5

Actual value meets specification.

Date: 2022-10-27


Niklas Lundqvist
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

