

Certificate of Analysis

Product Boule Micropipette Plastic EDTA 10x100 pcs
Boule Micropipette Plastic EDTA 1x100 pcs

Product No 1070030
1070039

Lot No 223645F1

Manufacturing Date 2022-10-10

Expiration Date 2025-09-28

Parameter	Specification
Volume 20 μ L	CV \leq 1,0%
WBC background	$\leq 0,2 \times 10^9/L$
RBC background	$\leq 0,03 \times 10^{12}/L$
HGB background	$\leq 0,2 \times \text{g/dL}$
PLT background	$\leq 10 \times 10^9/L$

Actual value meets specification.

Date: 2022-11-14



Sofia Karlström
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

