

Certificate of Analysis

Product SWELAB LUMI-L1 LYSE (200 mL)
Product No 1504515
Lot No 2022091601
Manufacturing Date 2022-09-16
Expiration Date 2024-09-15

Test	Specification
pH	2,0 – 2,4
CBC Performance test*	Passes test
Background WBC ($\times 10^9/L$)	$\leq 0,2$

*Performance testing at a reference laboratory. WBC, HGB and WBC scattergrams tested and approved.

Actual values meet specification.

Date: 2022-11-15



Pernilla Hammarstedt Rydberg
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

