

---

# Certificate of Analysis


**Product** EXIGO DILUENT  
**Product No** 1504328 (10 L)  
**Lot No** 2211-823  
**Manufacturing Date** 2022-11-21  
**Expiration Date** 2025-11-21

---

Test	Specification
pH	6.90 – 7.20
Osmolality (mOsm/kg)	300 – 330
Conductivity (mS/cm)	18.00 - 19.00
Particles (PLT/RBC channel)	≤ 5
Total bacterial growth (CFU/100ml)	≤ 5

Actual values meet specifications.

Date: 2022-11-30



*Pernilla Hammarstedt Rydberg*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

