


Certificate of Analysis

Product	Exigo System Cleaner, RFID (1.9 L)
Product No	1504503
Lot No	2212-831
Manufacturing Date	2022-12-05
Expiration Date	2025-12-05

Test	Specification
pH	7,00 – 8,00
Conductivity (mS/cm)	15,25 – 17,00
Particles (PLT/RBC channel)	≤ 30

Actual values meet specifications.

Date: 2022-12-13



Ulrika Miefalk
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

