

# Certificate of Analysis

**Product** Boule Micropipette Plastic EDTA 10x100 pcs  
Boule Micropipette Plastic EDTA 1x100 pcs  
**Product No** 1070030  
1070039  
**Lot No** 223645F3  
**Manufacturing Date** 2022-10-11  
**Expiration Date** 2025-09-28

---

Parameter	Specification
Volume 20 $\mu$ L	CV $\leq$ 1,0%
WBC background	$\leq$ 0,2 x 10 <sup>9</sup> /L
RBC background	$\leq$ 0,03 x 10 <sup>12</sup> /L
HGB background	$\leq$ 0,2 x g/dL
PLT background	$\leq$ 10 x 10 <sup>9</sup> /L

Actual value meets specification.

Date: 2022-11-22



*Niklas Lundqvist*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

