

# Certificate of Analysis

**Product** Boule Micropipette Plastic EDTA 10x100 pcs  
Boule Micropipette Plastic EDTA 1x100 pcs

**Product No** 1070030  
1070039

**Lot No** 224045N0

**Manufacturing Date** 2022-12-07

**Expiration Date** 2025-10-28

---

Parameter	Specification
Volume 20 $\mu$ L	CV $\leq$ 1,0%
WBC background	$\leq 0,2 \times 10^9/L$
RBC background	$\leq 0,03 \times 10^{12}/L$
HGB background	$\leq 0,2 \times g/dL$
PLT background	$\leq 10 \times 10^9/L$

Actual value meets specification.

Date: 2022-12-21



*Niklas Lundqvist*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

