

CERTIFICATE OF ANALYSIS

Product: Boule Cleaning Kit

Kit Lot No: 1762

Product Part #: 501-036


Kit Exp. Date: 2024-11-30

Part Number: 503-085			
Boule Detergent Cleaner Lot # 1763			
Manufacturing Date: 2022-11-07		Quality Testing Date: 2022-11-08	
Parameter	Specification	Result	Pass/Fail
pH (units):	7.60 – 8.40	8.33	Pass

Part Number: 503-086			
Boule Hypochlorite (2%) Cleaner Lot # 1764			
Manufacturing Date: 2022-11-03		Quality Testing Date: 2022-11-07	
Parameter	Specification	Result	Pass/Fail
pH (units):	11.80 – 13.00	12.99	Pass

Part Number: 503-087			
Boule Enzymatic Cleaner Lot # 1765			
Manufacturing Date: 2022-11-07		Quality Testing Date: 2022-11-08	
Parameter	Specification	Result	Pass/Fail
pH (units):	7.60 – 8.40	8.12	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.



Jeanne Roscoe
Quality and Regulatory Manager

12/7/2022
Date

