

CERTIFICATE OF ANALYSIS

Product: Swelab™ AlfaLyse

Lot No: 1773

Product Part #: 501-210

Exp. Date: 2025-12-31

Manufacturing Date: 01-18-2023

Quality Testing Date: 01-23-2023

Parameter	Specification	Result	Pass/Fail
pH (units):	7.54 – 8.30	8.24	Pass
Osmolality (mOsm/kg):	295 – 330	314	Pass
Conductivity (mS/cm):	18.00 – 18.50	18.34	Pass
Functional Test	Pass	Pass	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.


Quality Assurance

2/3/2023
Date

