

CERTIFICATE OF ANALYSIS

Product: Medonic™ M-Series Diluent

Lot No: 2212-001

Product Part #: 501-211


Exp. Date: 2025-11-30

Manufacturing Date: 2022-12-06

Quality Testing Date: 2022-12-08

Parameter	Specification	Result	Pass/Fail
pH (units):	6.90 – 7.20	7.19	Pass
Osmolality (mOsm/kg):	290 – 310	291	Pass
Conductivity (mS/cm):	17.50 – 18.00	17.53	Pass
Background Test (x10 ³ /μL):	0.0 – 5.0	0.25	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.



Jeanne Roscoe
Quality and Regulatory Manager

12/19/2022
Date

