

CERTIFICATE OF ANALYSIS

Product: Medonic™ M-Series Diluent

Lot No: 2301-004

Product Part #: 501-211

Exp. Date: 2025-12-31

Manufacturing Date: 2023-01-31

Quality Testing Date: 2023-02-03

Parameter	Specification	Result	Pass/Fail
pH (units):	6.90 – 7.20	7.11	Pass
Osmolality (mOsm/kg):	290 – 310	293	Pass
Conductivity (mS/cm):	17.50 – 18.00	17.57	Pass
Background Test (x10 ³ /μL):	0.0 – 5.0	0.3	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.



Quality and Regulatory

02-08-23

Date

