

CERTIFICATE OF ANALYSIS

Product: Quintus 5-Part Diluent

Lot No: 2301-201

Product Part #: 501-292

Exp. Date: 2024-12-31

Manufacturing Date: 2023-01-20

Quality Testing Date: 2023-02-10

Parameter	Specification	Result	Pass/Fail
pH (units):	6.7 – 7.1	7.1	Pass
Osmolality (mOsm/kg):	320 – 350	325	Pass
Conductivity (mS/cm):	18.50 – 19.50	18.68	Pass
Background (x10 ³ /μL)	0 – 5	3	Pass
Functional Test	Pass	Pass	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.



Quality & Regulatory Department

Date 2/16/23

