

# Certificate of Analysis

|                           |                                 |
|---------------------------|---------------------------------|
| <b>Product</b>            | SWELAB ALFADILUENT, RFID (20 L) |
| <b>Product No</b>         | 1504462                         |
| <b>Lot No</b>             | 2302-869                        |
| <b>Manufacturing Date</b> | 2023-02-13                      |
| <b>Expiration Date</b>    | 2026-02-13                      |

| <b>Test</b>                 | <b>Specification</b> |
|-----------------------------|----------------------|
| pH                          | 6,70 – 6,90          |
| Osmolality (mOsm/kg)        | 294 – 312            |
| Conductivity (mS/cm)        | 16,98 – 17,93        |
| Particles (PLT/RBC channel) | ≤ 5                  |

Actual value meets specification.

Date: 2023-02-20



Ulrika Miefalk  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

