

# Certificate of Analysis

**Product** Boule Micropipette Plastic EDTA 10x100 pcs  
Boule Micropipette Plastic EDTA 1x100 pcs

**Product No** 1070030  
1070039

**Lot No** 224745T4

**Manufacturing Date** 2023-01-09

**Expiration Date** 2025-11-28

---

Parameter	Specification
Volume 20 $\mu$ L	CV $\leq$ 1,0%
WBC background	$\leq 0,2 \times 10^9/L$
RBC background	$\leq 0,03 \times 10^{12}/L$
HGB background	$\leq 0,2 \times \text{g/dL}$
PLT background	$\leq 10 \times 10^9/L$

Actual value meets specification.

Date: 2023-02-14



Ulrika Miefalk  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

