

# Certificate of Analysis

**Product** BOULE CLEANING KIT (3 x 450 mL)  
**Product No** 1504111  
**Lot No** 2303-887  
**Manufacturing date** 2023-01-26  
**Use before** 2025-01-26

---

<b>Test Hypochlorite</b>	<b>Specification</b>
Hypochlorite concentration	2.0 – 2.4 %
Blank	PLT < 100 x 10 <sup>9</sup> /L RBC < 0.05 x 10 <sup>12</sup> /L

---

<b>Test Enzymatic Cleaner</b>	<b>Specification</b>
pH	7.6 – 8.4
Blank	PLT < 100 x 10 <sup>9</sup> /L RBC < 0.05 x 10 <sup>12</sup> /L


---

<b>Test Detergent Cleaner</b>	<b>Specification</b>
pH	7.6 – 8.4
Blank	PLT < 100 x 10 <sup>9</sup> /L RBC < 0.05 x 10 <sup>12</sup> /L

---

Actual values meet specifications.

Date: 2023-03-17



*Caroline Sparf*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

