

## CERTIFICATE OF ANALYSIS

**Product:** Medonic™ M-Series Lyse

**Lot No:** 2302-006

**Product Part #:** 501-212

**Exp. Date:** 2026-01-31

**Manufacturing Date:** 2023-02-24

**Quality Testing Date:** 2023-03-08

Parameter	Specification	Result	Pass/Fail
pH (units):	6.80 – 7.20	7.17	Pass
Osmolality (mOsm/kg):	220 – 240	230	Pass
Conductivity (mS/cm):	14.00 – 15.00	14.14	Pass
Functional Test	Pass	Pass	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.

  
Regulatory and Compliance

3/20/2023  
Date

