

Certificate of Analysis

Product Boule Micropipette Plastic EDTA 10x100 pcs
Boule Micropipette Plastic EDTA 1x100 pcs

Product No 1070030
1070039

Lot No 226045B5


Manufacturing Date 2023-02-28

Expiration Date 2026-01-28

Parameter	Specification
Volume 20 μ L	CV \leq 1,0%
WBC background	\leq 0,2 x 10 ⁹ /L
RBC background	\leq 0,03 x 10 ¹² /L
HGB background	\leq 0,2 x g/dL
PLT background	\leq 10 x 10 ⁹ /L

Actual value meets specification.

Date: 2023-03-20



Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

