

# Certificate of Analysis

**Product** Boule Micropipette Plastic EDTA 10x100 pcs  
Boule Micropipette Plastic EDTA 1x100 pcs

**Product No** 1070030  
1070039

**Lot No** 230145C9

**Manufacturing Date** 2023-03-07

**Expiration Date** 2026-02-28

---

Parameter	Specification
Volume 20 µL	CV ≤ 1,0%
WBC background	≤ 0,2 x 10 <sup>9</sup> /L
RBC background	≤ 0,03 x 10 <sup>12</sup> /L
HGB background	≤ 0,2 x g/dL
PLT background	≤ 10 x 10 <sup>9</sup> /L

Actual value meets specification.

Date: 2023-04-03



*Sofia Karlström*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

