

Certificate of Analysis

Product BOULE HYPOCHLORITE 4% CLEANER (100 mL)
Product No 1504385
Lot No 2304-901
Manufacturing date 2023-02-14
Use before 2024-08-31

Test Hypochlorite	Specification
Hypochlorite concentration	4 – 5 %
pH	11.5-13.5

Actual values meet specifications.

Date: 2023-04-20



Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

