

Certificate of Analysis

Product BOULE HYPOCHLORITE 2% CLEANER (500 mL)
Product No 1504113
Lot No 2304-900
Manufacturing date 2023-04-17
Expiration Date 2025-04-16

Test Hypochlorite	Specification
Hypochlorite concentration	2.0 – 2.4 %
pH	11.5 – 12.1
Blank	PLT < 100 x10 ⁹ /L RBC < 0.05 x 10 ¹² /L

Actual values meet specifications.

Date: 2023-05-05



Sofia Karlström
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

