

CERTIFICATE OF ANALYSIS

Product: Boule Cleaning Kit

Kit Lot No: 1822

Product Part #: 501-036

Kit Exp. Date: 2025-02-28

Part Number: 503-085			
Boule Detergent Cleaner Lot # 1819			
Manufacturing Date: 2023-02-27		Quality Testing Date: 2023-03-03	
Parameter	Specification	Result	Pass/Fail
pH (units):	7.60 – 8.40	8.22	Pass


Part Number: 503-086			
Boule Hypochlorite (2%) Cleaner Lot # 1820			
Manufacturing Date: 2023-02-23		Quality Testing Date: 2023-02-27	
Parameter	Specification	Result	Pass/Fail
pH (units):	11.80 – 13.00	12.96	Pass

Part Number: 503-087			
Boule Enzymatic Cleaner Lot # 1821			
Manufacturing Date: 2023-02-27		Quality Testing Date: 2023-03-03	
Parameter	Specification	Result	Pass/Fail
pH (units):	7.60 – 8.40	8.12	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.



Quality and Regulatory


Date 4/13/22 ^{EN} 4/13/23

