


Certificate of Analysis

Product	MEDONIC M51-D DILUENT (20L)
Product No	1504510
Lot No	2022122901
Manufacturing Date	2022-12-29
Expiration Date	2024-12-28

Test	Specification
pH	7,1 – 7,5
Osmolality (mOsm/kg)	290 – 310
Conductivity (mS/cm)	17,3 – 18,3
Particles PLT/RBC channel (x 10 ⁹ /L)	≤ 5

Actual values meet specification.

Date: 2023-05-22



Ulrika Miefalk
Ulrika Miefalk
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

