

# Certificate of Analysis

<b>Product</b>	MEDONIC M51-D DILUENT (20L)
<b>Product No</b>	1504510
<b>Lot No</b>	2023031302
<b>Manufacturing Date</b>	2023-03-13
<b>Expiration Date</b>	2025-03-12

<b>Test</b>	<b>Specification</b>
pH	7,1 – 7,5
Osmolality (mOsm/kg)	290 – 310
Conductivity (mS/cm)	17,3 – 18,3
Particles PLT/RBC channel (x 10 <sup>9</sup> /L)	≤ 5

Actual values meet specification.

Date: 2023-05-23



Ulrika Miefalk  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

