


Certificate of Analysis

Product	SWELAB ALFADILUENT (20 L)
Product No	1504124
Lot No	2305-916
Manufacturing Date	2023-05-22
Expiration Date	2026-05-21

Test	Specification
pH	6,70 – 6,90
Osmolality (mOsm/kg)	294 – 312
Conductivity (mS/cm)	16,98 – 17,93
Particles (PLT/RBC channel)	≤ 5

Actual value meets specification.

Date: 2023-05-24



Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

