

Boule Con-Diff

CONTROL

IVD



Control	LOT	
Tri-Level	22304-6K	08/31/23
Low	22304-61	08/31/23
Normal	22304-62	08/31/23
High	22304-63	09/04/23

Open vial stability 14 days¹

SWELAB

CONTROL	L
LOT	22304-61

Swelab Alfa/Alfa Plus

		Mean ²	± Range ³
RBC	10 ¹² /L	2.21	0.12
MCV	fL	65.1	5.0
RDW	%	14.8	4.0
Hct	%	14.4	2.5
Plt *	10 ⁹ /L	83	15
MPV	fL	11.4	1.5
WBC	10 ⁹ /L	3.4	0.3
Hgb	g/dl	5.3	0.3
MCH	pg	24.0	1.5
MCHC	g/dl	36.8	3.0
Lymph#	10 ⁹ /L	1.7	0.4
Mid#	10 ⁹ /L	0.1	0.3
Gran#	10 ⁹ /L	1.6	0.8
Lymph	%	48.9	7.0
Mid	%	3.5	5.0
Gran	%	47.6	7.0

CONTROL	N
LOT	22304-62

	Mean ²	± Range ³
	4.17	0.18
	86.8	5.0
	11.8	4.0
	36.2	3.0
	242	30
	11.5	1.5
	8.7	0.6
	12.6	0.4
	30.2	1.5
	34.8	3.0
	4.3	0.6
	0.3	0.5
	4.1	1.0
	48.9	5.0
	3.5	5.0
	47.6	5.0

CONTROL	H
LOT	22304-63

	Mean ²	± Range ³
	5.04	0.22
	94.6	5.0
	12.3	4.0
	47.7	3.5
	539	60
	11.9	1.5
	20.4	1.8
	16.3	0.5
	32.3	1.5
	34.2	3.0
	13.9	1.5
	0.6	1.0
	5.9	2.5
	68.0	5.0
	3.0	5.0
	29.0	5.0

- 1
- bg Трайност на отворен флакон 14 дни (bg)
 - hr Stabilnost otvorene bočice 14 dani (hr)
 - cs Stabilita otevřené ampulky 14 dní (cs)
 - da Åben r r stabilitet 14 dage (da)
 - et Stabiilsus avatud viaalis 14 p evade (et)
 - fr Stabilit  en flacon ouvert 14 jour (fr)
 - de Stabilit t ge ffneter Flaschen 14 tage (de)
 - el Διατήρηση δείγματος μετά την αποσφράγιση 14 ημε ρα (el)
 - it Stabilit  della fiala aperta 14 giorni (it)
 - lv Atv rt flakonu stabilit ti 14 dienas (lv)
 - lt Stabilumas atidarius buteliuk  14 dienas (lt)
 - no Åpen r r/glass stabilitet 14 dager (no)
 - pl Trwa o c po otworzeniu fiołki 14 Liczba dni (pl)
 - pt Estabilidade ap s abertura do frasco 14 dias (pt)
 - ro Termenul de valabilitate al fiolei desf cute 14 zile (ro)
 - ru Стабильность вскрытого флакона 14 дней (ru)
 - sr Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
 - sk Stabilita otvorenej injek nej liekovky 14 dní (sk)
 - sl Stabilita otvorenej injek nej liekovky 14 dni (sl)
 - es Estabilidad de la c psula abierta 14 dias (es)
 - sv H llbarhet f r  ppnad flaska 14 dagar (sv)
 - tr A ık ŐiŐe stabilitesi 14 g n (tr)

- 2
- Средно
 - Sredstvo
 - Pr m r
 - Middelv rdi
 - Keskmine
 - Moyenne
 - Mittelwert
 - Μ ση
 - Media
 - Noz m 
 - Vidurkis
 - Middelverdi
 - Srednia
 - M dia
 - Medie
 - Среднее
 - Srednja vrednost
 - Priemer
 - Povpre no
 - Media
 - Medelv rde
 - Ortalama

- 3
- Обхват
 - Raspon
 - Rozsah
 - Omr de
 - Vahemik
 - Intervalle
 - Bereich
 - Αναμενόμενο ε ρος α
 - Intervallo
 - Diapazons
 - Intervalas
 - Omr de
 - Zakres
 - Intervalo
 - Interval
 - диапазон
 - Opseg
 - Rozsah
 - Obmo je
 - Intervalo
 - Intervall
 - Aralıđı



[http://www.swelab.com/
support/downloads](http://www.swelab.com/support/downloads)



1504054

For Swelab Alfa / Alfa Plus with barcode reader

Alfa:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Alfa Plus:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: Swelab Alfa
Lot: 22304-61
Prod.: Con-Diff Low
ID: 2230461+
Exp.: 31-Aug-2023

Instr.: Swelab Alfa
Lot: 22304-62
Prod.: Con-Diff Norm
ID: 2230462+
Exp.: 31-Aug-2023

Instr.: Swelab Alfa
Lot: 22304-63
Prod.: Con-Diff High
ID: 2230463+
Exp.: 4-Sep-2023

