

Certificate of Analysis


Product SWELAB LUMI-L1 LYSE (200 mL)
Product No 1504515
Lot No 2023040101
Manufacturing Date 2023-04-01
Expiration Date 2025-03-31

Test	Specification
pH	2,0 – 2,4
CBC Performance test*	Passes test
Background WBC (x 10 ⁹ /L)	≤ 0,2

*Performance testing at a reference laboratory. WBC and HGB tested and approved.

Actual values meet specification.

Date: 2023-06-27



Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

