a Boule Diagnostics, AB company



## **CERTIFICATE OF ANALYSIS**

Product: Quintus 5-Part Stopper

**Lot No:** 2103-202

**Product Part #: 501-350** 

**Exp. Date:** 2024-02-29

Manufacturing Date: 2021-03-29

**Quality Testing Date: 2021-04-07** 

Parameter	Specification	Result	Pass/Fail
pH (units):	10.10 – 11.20	11.18	Pass
Osmolality (mOsm/kg):	830 – 860	839	Pass
Conductivity (mS/cm):	42.7 – 43.3	43.0	Pass
Functional Test	Pass	Pass	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.

Jeanne Roscoe

Quality and Regulatory Manager

npn i o j a

SEAL 1998