

CERTIFICATE OF ANALYSIS

Product: Medonic™ M-Series Lyse

Lot No: 2303-002

Product Part #: 501-212

Exp. Date: 2026-02-28

Manufacturing Date: 2023-03-24

Quality Testing Date: 2023-04-13

Parameter	Specification	Result	Pass/Fail
pH (units):	6.80 – 7.20	7.13	Pass
Osmolality (mOsm/kg):	220 – 240	232	Pass
Conductivity (mS/cm):	14.00 – 15.00	14.18	Pass
Functional Test	Pass	Pass	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.


Regulatory and Compliance

4/24/23
Date

