

# CERTIFICATE OF ANALYSIS

**Product:** Boule Cleaning Kit

**Kit Lot No:** 1863

**Product Part #:** 501-036

**Kit Exp. Date:** 2025-05-31

<b>Part Number:</b> 503-085			
<b>Boule Detergent Cleaner Lot #</b> 1860			
<b>Manufacturing Date:</b> 2023-06-08		<b>Quality Testing Date:</b> 2023-06-02	
Parameter	Specification	Result	Pass/Fail
pH (units):	7.60 – 8.40	8.25	Pass

<b>Part Number:</b> 503-086			
<b>Boule Hypochlorite (2%) Cleaner Lot #</b> 1861			
<b>Manufacturing Date:</b> 2023-06-08		<b>Quality Testing Date:</b> 2023-06-01	
Parameter	Specification	Result	Pass/Fail
pH (units):	11.80 – 13.00	12.93	Pass

<b>Part Number:</b> 503-087			
<b>Boule Enzymatic Cleaner Lot #</b> 1874			
<b>Manufacturing Date:</b> 2023-06-08		<b>Quality Testing Date:</b> 2023-06-09	
Parameter	Specification	Result	Pass/Fail
pH (units):	7.60 – 8.40	8.18	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.



Quality and Regulatory

07/14/23

Date

