

CERTIFICATE OF ANALYSIS

Product: Swelab™AlfaDiluent

Lot No: 2307-002

Product Part #: 501-208


Exp. Date: 2026-06-30

Manufacturing Date: 07-07-2023

Quality Testing Date: 07-14-2023

Parameter	Specification	Result	Pass/Fail
pH (units):	6.50 – 6.90	6.83	Pass
Osmolality (mOsm/kg):	294 – 312	299	Pass
Conductivity (mS/cm):	16.98 – 17.93	17.19	Pass
Background (x10 ³ /μL):	0.0 – 8.0	0.8	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.


Quality Assurance

07/19/23
Date

