

CERTIFICATE OF ANALYSIS

Product: Swelab™ AlfaLyse 5L RFID

Lot No: 2307-003

Product Part #: 501-466


Exp. Date: 2026-06-30

Manufacturing Date: 2023-07-18

Quality Testing Date: 2023-07-24

Parameter	Specification	Result	Pass/Fail
pH (units):	7.54 – 8.30	8.26	Pass
Osmolality (mOsm/kg):	295 – 330	319	Pass
Conductivity (mS/cm):	18.00 – 18.50	18.26	Pass
Functional Test	Pass/Fail	Pass	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.


Quality Assurance

07/26/23
Date

