

Certificate of Analysis

Product H50V LYSE, RFID
Product No 1440109 (500 mL)
Lot No 23051351M7

Manufacturing Date 2023-06-12

Expiration Date 2025-06-11

Test	Specification
Appearance	Clear liquid, no precipitate particles or flocs
pH	2.00 – 4.50
Absorbance λ_{\max} (nm)	530 – 550
Blank	WBC $\leq 0.3 \times 10^9/L$ HGB $\leq 2g/L$
WBC	a) Has two population peaks b) Meet the peak form and position mark requirements.

Actual values meet specifications.

Date: 2023-07-18



Pernilla Hammarstedt Rydberg
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

